Title:	**Drivers for STR240**		ID:
			0331
Date in:	**Response:**	**Model:**	**Author:**
2008-10-16	*2008-10-16*	*STR240*	*CMa*

Q:

Liconic STR240-HCIT
Cat # 9143 00 92
Is it possible to use the usual Driver (STX Server)?

A:

Yes. All our Units are controlled with the identical commands.