

Title:	Retrofit of CO2	ID:	
		0175	
Date in:	Response:	Model:	Author:
2003-10-23	2003-10-23	-	CMa

Q:

Is it possible to retrofit or exchange our Liconic 20-50°C Incubator with CO2 option?
How much it would be?
And how soon it can be done; we would like to have some kind of solution in a month or so time frame.
And how about Shaking option?

A:

The options can be retrofit. The retrofit of CO2 and/or shaker involves some changes that require this work to be done at the manufacturer. This means the unit would have to come back for retrofit. If everything can be prepared the retrofit can be done within one week.